

Labnet Enduro™ MiniMix Nutating Mixer

Instruction Manual

Catalog Numbers:

S0600

S0600-230V

S0600-230V-UK



Table of Contents

1.0 Introduction	1
2.0 Symbols and Conventions.....	1
3.0 Technical Specifications.....	1
4.0 Installation.....	2
5.0 Operating Instructions.....	2
6.0 Cleaning and Maintenance	2
7.0 Troubleshooting.....	2
8.0 Limited Warranty.....	2
9.0 Equipment Disposal	3

1.0 Introduction

The Labnet Enduro™ MiniMix is a three dimensional nutating mixer typically used for maintaining the uniform suspension of liquids over transfer membranes with a minimum volume. The Enduro MiniMix is also ideal for processing Western Blots and staining and destaining PAGE gels.

The Enduro MiniMix achieves its 3-dimensional mixing by combining the rotating motion of an orbital shaker with the rocking motion of a platform rocker. The modern styling of the Enduro MiniMix's housing is achieved through the use of injection molded FRABS plastic. The small platform is constructed of powder coated aluminum topped with a rubber TPE mat.

2.0 Symbols and Conventions



The electrical warning indicates the presence of a potential hazard which could result in electrical shock.



CAUTION: This symbol refers you to important operating and maintenance (servicing) instructions within this Instruction Manual. Failure to heed this information may present a risk of damage or injury to persons or equipment.



This symbol identifies a Protective Earth (PE) terminal, which is provided for connection of the supply system's protective earth (green or green/yellow) conductor.

3.0 Technical Specifications

Motion	3-dimensional rocking rotating
Speed	
S0600	18 rpm
S0600-230V	20 rpm
Platform Tilt Angle	5°, fixed
Dimensions (W x D x H)	9.5 x 5.75 x 6.75 in. (24 x 14.6 x 17 cm)
Platform Dimensions	9.5 x 6.75 in. (20 x 16.5 cm)
Weight	1.88 lbs. (0.85 kg)
Platform Load Capacity	1.75 lbs. (0.8 kg), balanced
Platform Capacity without Stacking	Four 8 x 10 cm (Bio-Rad, NuSep) or two 10 x 10 cm (Lonza, Invitrogen) blotting boxes
Electrical	
S0600	120V~, 50/60Hz, 0.03A, 3.6W
S0600-230V	230V~, 50/60Hz, 0.017A, 4W
Temperature Range	5°C to 40°C (up to 80% RH at 31°C, decreasing linearly to 50% RH at 40°C)
Protection Class	IP21

The Labnet Enduro MiniMix is designed to be safe at least when operated under the following conditions:

- ▶ Indoor use only
- ▶ Altitude up to 2,000 meters
- ▶ Pollution degree 2

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

NOTE: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. The equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the user manual, may cause harmful interference to radio communications. Operation of equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference.



CAUTION: This equipment is not intended for use in residential environments and may not provide adequate protection to radio reception in such environments.

This ISM device complies with Canadian ICES-001.

Cet appareil ISM est conforme à la norme NMB-001 du Canada.

4.0 Installation

The Labnet Enduro™ MiniMix can be used on a bench, in a cold room or in an incubator up to 40°C. Place the unit on a flat stable surface and be sure to leave sufficient room around the unit to allow both the platform and load to move freely. A minimum of 4 inches (10 cm) clearance on all sides of the mixer is advised.

Plug the unit into a properly grounded outlet of the appropriate voltage as shown on the unit's data plate label.

5.0 Operating Instructions

The flat matted pattern is configured to allow four 8 x 10 cm (Bio-Rad, NuSep) or two 10 x 10 cm (Lonza, Invitrogen) blotting boxes to be securely placed on the mat without additional strapping. If blotting boxes are to be stacked, they should be secured with rubber bands. Holes in the corners of the platform edges are provided for this purpose.

Always load the platform in a uniform and balanced fashion.

Use the power switch to turn the unit on. The direction of rotation in which the unit starts is random and does not affect mixing. If desired, the rotational direction can be reversed by holding the unit's base and very briefly touching the rotating platform on the bottom at one corner. This technique can also be used to select the direction of rotation.

6.0 Cleaning and Maintenance

Always disconnect the cord from the power source before maintaining or servicing.

The Labnet Enduro MiniMix requires no maintenance other than occasional cleaning with a damp cloth with mild detergent. The rubber mat may be washed in soap and water. **NOTE:** Do not immerse the unit in water or spill liquids over the unit. Be sure the unit is dry before reconnecting to power.

7.0 Troubleshooting

Should you have a question about the operation of the Labnet Enduro MiniMix or if service is required, contact Corning Customer Service. Do not send in a unit for service without first calling to obtain a repair authorization number. Should the unit require return for service, it should be properly packed to avoid damage. Any damage resulting from improper packaging shall be the responsibility of the user.

8.0 Limited Warranty

Corning Incorporated (Corning) warrants that this product will be free from defects in material and workmanship for a period of one (1) year from date of purchase. CORNING DISCLAIMS ALL OTHER WARRANTIES WHETHER EXPRESSED OR IMPLIED, INCLUDING ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR OF FITNESS FOR A PARTICULAR PURPOSE. Corning's sole obligation shall be to repair or replace, at its option, any product or part thereof that proves defective in material or workmanship within the warranty period, provided the purchaser notifies Corning of any such defect. Corning is not liable for any incidental or consequential damages, commercial loss or any other damages from the use of this product.

This warranty is valid only if the product is used for its intended purpose and within the guidelines specified in the supplied instruction manual. This warranty does not cover damage caused by accident, neglect, misuse, improper service, natural forces or other causes not arising from defects in original material or workmanship. This warranty does not cover motor brushes, fuses, light bulbs, batteries or damage to paint or finish. Claims for transit damage should be filed with the transportation carrier.

In the event this product fails within the specified period of time because of a defect in material or workmanship, contact Corning Customer Service at: USA/Canada 1.800.492.1110, outside the U.S. +1.978.442.2200, visit www.corning.com/lifesciences, or contact your local support office.

Corning's Customer Service team will help arrange local service where available or coordinate a return authorization number and shipping instructions. Products received without proper authorization will be returned. All items returned for service should be sent postage prepaid in the original packaging or other suitable carton, padded to avoid damage. Corning will not be responsible for damage incurred by improper packaging. Corning may elect for onsite service for larger equipment.

Some states do not allow limitation on the length of implied warranties or the exclusion or limitation of incidental or consequential damages. This warranty gives you specific legal rights. You may have other rights which vary from state to state.

No individual may accept for, or on behalf of Corning, any other obligation of liability, or extend the period of this warranty.

For your reference, make a note of the serial and model number, date of purchase, and supplier here.

Serial No. _____ Date Purchased _____

Model No. _____ Supplier _____

9.0 Equipment Disposal



According to Directive 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE), this product is marked with the crossed-out wheeled bin and must not be disposed of with domestic waste.

Consequently, the buyer shall follow the instructions for reuse and recycling of waste electronic and electrical equipment (WEEE) provided with the products and available at www.corning.com/weee.

To request certificates, please contact us at www.labnetlink.com.

Warranty/Disclaimer: Unless otherwise specified, all products are for research use or general laboratory use only.* Not intended for use in diagnostic or therapeutic procedures. Not for use in humans. These products are not intended to mitigate the presence of microorganisms on surfaces or in the environment, where such organisms can be deleterious to humans or the environment. Corning Life Sciences makes no claims regarding the performance of these products for clinical or diagnostic applications. *For a listing of US medical devices, regulatory classifications or specific information on claims, visit www.corning.com/resources.

Corning's products are not specifically designed and tested for diagnostic testing. Many Corning products, though not specific for diagnostic testing, can be used in the workflow and preparation of the test at the customers discretion. Customers may use these products to support their claims. We cannot make any claims or statements that our products are approved for diagnostic testing either directly or indirectly. The customer is responsible for any testing, validation, and/or regulatory submissions that may be required to support the safety and efficacy of their intended application.

CORNING

Corning Incorporated
Life Sciences

www.corning.com/lifesciences

NORTH AMERICA

t 800.492.1110
t 978.442.2200

ASIA/PACIFIC

Australia/New Zealand
t 61 427286832
Chinese Mainland
t 86 21 3338 4338

India

t 91 124 4604000

Japan

t 81 3-3586 1996

Korea

t 82 2-796-9500

Singapore

t 65 6572-9740

Taiwan

t 886 2-2716-0338

EUROPE

CEurope@corning.com

LATIN AMERICA

grupoLA@corning.com

Brazil

t 55 (11) 3089-7400

Mexico

t (52-81) 8158-8400